


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 1119 Accredited to ISO/IEC 17025:2017	<b>London Borough of Enfield</b>  <b>Issue No: 018    Issue date: 19 June 2024</b>	
	<b>Asbestos Team</b> Civic Centre Silver Street Enfield Middlesex EN1 3ES	<b>Contact: Mr M A Gifford</b> <b>Tel: +44 (0)20 3821 1798</b> <b>E-Mail: asbestos@enfield.gov.uk</b> <b>Website: www.enfield.gov.uk</b>

**Testing performed by the Organisation at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<b>Address</b> Asbestos Team Civic Centre Silver Street Enfield Middlesex EN1 3ES	<b>Local contact</b> Mr M A Gifford  Tel: +44 20 3821 1798  Email: asbestos@enfield.gov.uk Website: www.enfield.gov.uk	Health and Hygiene Head Office  Asbestos – All Support Functions	A

#### Site activities performed away from the locations listed above:

Location details		Activity	Location code
Client Premises	As above	Health and Hygiene	B



1119  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**London Borough of Enfield**  
**Issue No:** 018    **Issue date:** 19 June 2024

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method – Quality Manual, Section 9 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method – Quality Manual, Section 9 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			