


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>1366</b> Accredited to <b>ISO/IEC 17025:2017</b>	<b>ITW Ltd trading as Avery Weigh-Tronix (Calibration Select Laboratory)</b>	
	Issue No: 031    Issue date: 24 August 2021	
	<b>Calibration Select</b> Foundry Lane Smethwick West Midlands B66 2LP	<b>Contact: Mr M Moran</b> Tel: +44 (0) 121 568 1407 Fax: +44 (0) 121 697 1407 E-Mail: mmoran@awtx-itw.com Website: www.calibrationselect.co.uk
<b>Testing performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
LOAD CELLS UP TO 50 TONNE CAPACITY	Determination of load cell error, repeatability error and temperature effect on minimum dead load output  Determination of creep error  Determination of minimum dead load output return  Determination of barometric pressure effects  Determination of humidity effects  Additional tests for load cells equipped with electronics	OIML R60 Edition 2000 (E) Accuracy Class C and D up to a number of verification intervals consistent with the measurement capabilities.
END		