


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS TESTING 1696 Accredited to ISO/IEC 17025:2017	West Yorkshire Analytical Services a part of West Yorkshire Joint Services	
	Issue No: 048 Issue date: 28 August 2020	
	Nepshaw Lane South Morley Leeds LS27 7JQ	Contact: Service Manager / Quality Manager Tel: +44 (0)113 535 0177 E-Mail : asbestos@wyjs.org.uk Website: https://www.wyjs.org.uk/analytical
Testing performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Nepshaw Lane South Morley Leeds LS27 7JQ	Local contact Contact: Service Manager / Quality Manager Tel: +44 (0)113 535 0177 E-Mail: https://www.wyjs.org.uk/analytical	Head Office Asbestos A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Asbestos	B
Mobile laboratory (Registration DC17 TCW)	Asbestos	C



1696

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

West Yorkshire Analytical Services
a part of West Yorkshire Joint Services
Issue No: 048 Issue date: 28 August 2020

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre Counting	HSG 248: February 2005 Documented In-House procedure METH1009	A, B, C
	Sampling of air for fibre counting	HSG 248: February 2005 Documented In-House procedure METH1009	B, C
	4 Stage Clearance Process	HSG 248: February 2005 Documented In-House procedure METH1009	B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248: February 2005 by Documented In-House Procedure 1003 using stereo-microscopy polarised light microscopy and dispersion staining	A, C
END			