


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>1709</b> Accredited to <b>ISO/IEC 17025:2017</b>	<b>RPS Consulting Services Ltd</b> <b>trading as RPS Consultants</b> <b>Issue No: 075 Issue date: 08 January 2021</b>	
	<b>20 Farringdon Street</b> London <b>EC4A 4AB</b>	<b>Contact: Mr A Craig</b> <b>Tel: +44 (0) 20 3691 0500</b> <b>E-Mail: andrew.craig@rpsgroup.com</b> <b>Website: www.rpsgroup.com</b>
<b>Testing performed by the Organisation at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> 20 Farringdon Street London EC4A 4AB United Kingdom	Health and Hygiene Support Functions: Quality Management, including contract review, document control, auditing and quality control	D
<b>Local contact</b> Mr A Craig Tel: +44 (0)20 3691 0500 Email: andrew.craig@rpsgroup.com	Legionella Sampling Support Functions: Quality Management, including contract review, document control, auditing and quality control	D
<b>Local contact</b> Mr M Rose Tel: +44 (0)20 3691 0500 Email: rpsleg@rpsgroup.com		

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene Environmental and Water Sampling and Testing	London - K



1709  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**RPS Consulting Services Ltd**  
trading as **RPS Consultants**  
Issue No: 075 Issue date: 08 January 2021

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	K
	Fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	D, K
	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)	K
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Sampling of bulk materials for asbestos identification	HSG 248:February 2005 (Documented In-House Procedure)	K



1709  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**RPS Consulting Services Ltd**  
trading as **RPS Consultants**  
Issue No: 075 Issue date: 08 January 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>WATERS (non regulatory) Drinking Water (mains), Process Water (including Cooling Towers), Recreational Water including Pool Water</p> <p>From sampling location including:-</p> <p>Cooling Towers, Domestic Hot &amp; Cold water systems. Calorifiers, Water dispensers, Cold water tanks, Humidifiers / water, Spa Pools, TMV/mixer taps, Instantaneous water heater, Showers/spray taps and Circulating condenser water (closed open circuit towers)</p>	<p>Water sampling for the purpose of facilities management and as a stand alone practice, for subsequent microbiological analysis:</p> <ul style="list-style-type: none"> <li>• Aerobic Colony Count</li> <li>• Legionella</li> </ul>	<p>Documented in house method RPSCL/1 based on BS 7592:2008, BS 8554:2015 and The Microbiology of Drinking Water part 2 2010</p>	K
END			