


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 1709 Accredited to ISO/IEC 17025:2017	RPS Consulting Services Ltd t/a RPS Consultants Issue No: 079 Issue date: 06 March 2025	
	20 Farringdon Street London EC4A 4AB	Contact: Mr M Rose Tel: +44 (0) 20 3691 0500 E-Mail: mike.rose1@rps.tetrattech.com Website: www.rpsgroup.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 20 Farringdon Street London EC4A 4AB United Kingdom Local contact Mr M Rose Tel: +44 (0)20 3691 0500 Email: mike.rose1@rps.tetrattech.com Mr M Rose Tel: +44 (0)20 3691 0500 Email: mike.rose1@rps.tetrattech.com	Health and Hygiene Head Office	D
	Asbestos – All Support Functions Legionella Sampling Support Functions: Quality Management, including contract review, document control, auditing and quality control	D

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene Environmental and Water Sampling and Testing	London - K
Temporary Site First Floor Vanguard House Vanguard Way Hounslow TW6 2FH	Health and Hygiene Environmental Sampling and Testing	L



1709
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RPS Consulting Services Ltd
t/a RPS Consultants
Issue No: 079 Issue date: 06 March 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method RPSCA1 based on HSG 248	K, L
	Fibre counting	Documented In-House Method RPSCA1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	D, K, L
	4 Stage Clearance Process	Documented In-House Method RPSCA1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	K
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method RPSCA2 based on HSG 248	K



1709
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RPS Consulting Services Ltd
t/a RPS Consultants
Issue No: 079 Issue date: 06 March 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>WATERS (non regulatory) Drinking Water (mains), Process Water (including Cooling Towers), Recreational Water including Pool Water</p> <p>From sampling location including:-</p> <p>Cooling Towers, Domestic Hot & Cold water systems. Calorifiers, Water dispensers, Cold water tanks, Humidifiers / water, Spa Pools, TMV/mixer taps, Instantaneous water heater, Showers/spray taps and Circulating condenser water (closed open circuit towers)</p>	<p>Water sampling for the purpose of facilities management and as a stand alone practice, for subsequent microbiological analysis:</p> <ul style="list-style-type: none"> • Aerobic Colony Count • <i>Legionella</i> spp. • Coliforms • <i>Escherichia coli</i> • <i>Pseudomonas aeruginosa</i> 	<p>Documented in house method RPSCL/1 based on BS 7592:2022, BS 8554:2015 and The Microbiology of Drinking Water part 2 2010</p>	K
END			