


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 1752 Accredited to ISO/IEC 17025:2017	Tersus Consultancy Ltd	
	Issue No: 055	Issue date: 01 December 2024
	1st Floor Rainham House Manor Way Rainham Essex RM13 8RH	Contact: Mr C Simmonite Tel: +44 (0) 1909 560673 E-Mail: craig.simmonite@tersusgroup.co.uk Website: www.tersusgroup.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 1 st Floor Rainham House Manor Way Rainham Essex RM13 8RH Local contact Mr T Rose Tel: +44 (0)1708 555705 Email: Trevor.Rose@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Head Office Asbestos – All Support Functions	A
Address Unit 1 Manor Drive Outgang Lane Dinnington Sheffield S25 3QU Local contact Mrs G Purewal Tel: +44 (0)1909 560673 Email: Gurpreet.Purewal@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	B
Address Red Mill House Red Mill Estate Wednesbury WS10 0NN Local contact Mrs G Purewal Tel: +44 (0)121 244 1828 Email: Gurpreet.Purewal@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	C



1752

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tersus Consultancy Ltd
Issue No: 055 Issue date: 01 December 2024

Testing performed by the Organisation at the locations specified

Location details		Activity	Location code
Address Unit 1, Medipark Mallard Way Strathclyde Business Park Bellshill ML4 3BF	Local contact Mrs G Purewal Tel: +44 (0)1909 560673 Email: Gurpreet.Purewal@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	D



1752

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tersus Consultancy Ltd
Issue No: 055 Issue date: 01 December 2024

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
Address Axys House Parc Nantgarw Cardiff CF15 7TW Local contact Mr C Jones Tel: +44 (0) 1278 654 345 Email: Craig.Jones@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	G
Address Health Physics R S R L Building D60 Winfrith Dorchester DT2 8WQ Local contact Mr D Evans Tel: +44 (0) 1278 654 345 Email: David.Evans@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	H
Address 10 Rutherford Avenue Didcot OX11 0DF Local contact Mr D Evans Tel: +44 (0) 1278 654 345 Email: David.Evans@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	I



1752

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tersus Consultancy Ltd
Issue No: 055 Issue date: 01 December 2024

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
Address Hinkley Point A Power Station Bridgewater TE5 1YA Local contact Mr D Evans Tel: +44 (0) 1278 654 345 Email: David.Evans@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	J
Address Dungeness A Power Station Romney Marsh TN29 9PP Local contact Mr D Evans Tel: +44 (0) 1278 654 345 Email: David.Evans@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	K
Address Oldbury Power Station Oldbury Naite Thornbury BS35 1RQ Local contact Mr D Evans Tel: +44 (0) 7914 668960 Email: David.Evans@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	L



1752

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tersus Consultancy Ltd
Issue No: 055 Issue date: 01 December 2024

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
Address Berkeley Power Station Berkerley GL13 9PA Local contact Mr D Evans Tel: +44 (0) 7921 511 417 Email: David.Evans@tersusgroup.co.uk Website: www.tersusgroup.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	M

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including: <ul style="list-style-type: none">- Domestic- Commercial- Industrial	Health and Hygiene	N
Mobile Testing Laboratories – Asbestos Fibres in Air	Health and Hygiene	Sheffield – O Rainham - P



1752

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tersus Consultancy Ltd
Issue No: 055 Issue date: 01 December 2024

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Flexible scope - to establish new temporary site laboratories to conduct the following asbestos testing activities: <ul style="list-style-type: none">• <i>Fibre counting</i>• <i>Identification of asbestos in bulk material</i> The Site laboratories (denoted *) are set up in accordance with the Documented In-House Procedure IMSP69 Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Sampling of air for fibre counting	Documented In-House Method GDP4 based on HSG 248	A, B, G, H*, I*, J*, K*, L*, M*, N, O, P
	Fibre counting	Documented In-House Method GDP4, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, G, H*, I*, J*, K*, L*, M*, N, O, P
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method GDP4, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	H*, I*, J*, K*, L*, M*, N
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method GDP1 based on HSG 248	H*, I*, J*, K*, L*, M*, N
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method GDP3 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, B, C, D, G, I*, J*, K*, L*, M*,
END			