


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING</p> <p>1752</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p><b>Tersus Consultancy Ltd</b></p> <p>Issue No: 047 Issue date: 30 September 2022</p>	
	<p>1st Floor Rainham House Manor Way Rainham Essex RM13 8RH</p>	<p>Contact: Ms K Hukin Tel: +44 (0) 1909 560673 E-Mail: kylie.hukin@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>
<p>Testing performed by the Organisation at the locations specified below</p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b> 1<sup>st</sup> Floor Rainham House Manor Way Rainham Essex RM13 8RH</p> <p><b>Local contact</b> Mr T Rose  Tel: +44 (0)1708 555705 Email: trevor.rose@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Head Office</p> <p>Asbestos – All Support Functions</p>	A
<p><b>Address</b> Unit 6 Carrera Court Dinnington Sheffield S25 2RG</p> <p><b>Local contact</b> Mr S Morley  Tel: +44 (0)1909 560673 Email: simon.morley@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	B



1752

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Tersus Consultancy Ltd**  
Issue No: 047 Issue date:30 September 2022

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
<p><b>Address</b> Red Mill House Red Mill Estate Wednesbury WS10 0NN</p> <p><b>Local contact</b> Miss G Cheema  Tel: +44 (0)121 244 1828 Email: gurpreet.cheema@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"><li>• <i>Quality Audit</i></li><li>• <i>Administration</i></li><li>• <i>Contract Review</i></li><li>• <i>Scheduling</i></li><li>• <i>Personnel</i></li><li>• <i>Equipment</i></li><li>• <i>Measurement</i></li><li>• <i>Traceability</i></li><li>• <i>Reporting</i></li></ul>	C
<p><b>Address</b> 1325 Park Avenue, Aztec West, Bristol BS32 4RX</p> <p><b>Local contact</b> Mr C Horton  Tel: +44 (0) 1173 702506 Email: chris.horton@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"><li>• <i>Quality Audit</i></li><li>• <i>Administration</i></li><li>• <i>Contract Review</i></li><li>• <i>Scheduling</i></li><li>• <i>Personnel</i></li><li>• <i>Equipment</i></li><li>• <i>Measurement</i></li><li>• <i>Traceability</i></li><li>• <i>Reporting</i></li></ul>	D



1752

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Tersus Consultancy Ltd**  
**Issue No: 047 Issue date:30 September 2022**

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
<p><b>Address</b> Unit C6 Dolphin Enterprise Centre Evershed Way Shoreham-by-Sea West Sussex BN43 6QB</p> <p><b>Local contact</b> Mr J Senior  Tel: +44 (0)1273 621100 Email: josh.senior@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	E
<p><b>Address</b> Eurosafes House Tribune Way Centurion Park Clifton Moor York YO30 4RY</p> <p><b>Local Contact</b> Ms J Ormshaw  Tel: -44 (0) 1904 691515 Email: jen.ormshaw@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Reporting</i></li> </ul>	F
<p><b>Address</b> Axys House Parc Nantgarw Cardiff CF15 7TW</p> <p><b>Local contact</b> Mr D Evans  Tel: +44 (0) 1278 654 345 Email: David.Evans@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	G



1752

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Tersus Consultancy Ltd**

**Issue No: 047 Issue date:30 September 2022**

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
<p><b>Address</b> Health Physics R S R L Building D60 Winfrith Dorchester DT2 8WQ</p> <p><b>Local contact</b> Mr D Wasilewski  Tel: +44 (0) 1278 654 345 Email: David.wasilewski@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	H
<p><b>Address</b> 10 Rutherford Avenue Didcot OX11 0DF</p> <p><b>Local contact</b> Mr D Wasilewski  Tel: +44 (0) 1278 654 345 Email: David.wasilewski@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	I
<p><b>Address</b> Hinkley Point A Power Station Bridgewater TE5 1YA</p> <p><b>Local contact</b> Mr D Wasilewski  Tel: +44 (0) 1278 654 345 Email: David.wasilewski@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	J



1752

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Tersus Consultancy Ltd**  
**Issue No: 047 Issue date:30 September 2022**

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
<p><b>Address</b> Dungeness A Power Station Romney Marsh TN29 9PP</p> <p><b>Local contact</b> Mr D Wasilewski  Tel: +44 (0) 1278 654 345 Email: David.wasilewski@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	<p>K</p>
<p><b>Address</b> Oldbury Power Station Oldbury Naite Thornbury BS35 1RQ</p> <p><b>Local contact</b> Mr J Bolt  Tel: +44 (0) 7914 668960 Email: james.bolt@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	<p>L</p>
<p><b>Address</b> Berkeley Power Station Berkerley G13 9PA</p> <p><b>Local contact</b> Mr D Rolfe  Tel: +44 (0) 7921 511 417 Email: david.rolfe@tersusgroup.co.uk Website: www.tersusgroup.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	<p>M</p>



1752

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Tersus Consultancy Ltd**  
**Issue No: 047 Issue date:30 September 2022**

Testing performed by the Organisation at the locations specified

**Site activities performed away from the locations listed above:**

Location details	Activity	Location code
Premises including: - Domestic - Commercial - Industrial	Health and Hygiene	N
Mobile Testing Laboratories – Asbestos Fibres in Air	Health and Hygiene	Sheffield – O Shoreham-by-Sea - P
Mobile Testing Laboratories – Asbestos in Bulk Materials	Health and Hygiene	Sheffield - Q



1752

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Tersus Consultancy Ltd

Issue No: 047 Issue date: 30 September 2022

Testing performed by the Organisation at the locations specified

#### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Flexible scope - to establish new temporary site laboratories to conduct the following asbestos testing activities: <ul style="list-style-type: none"> <li>Fibre counting</li> <li>Identification of asbestos in bulk material</li> </ul> The Site laboratories (denoted *) are set up in accordance with the Documented In-House Procedure IMSP69  Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Sampling of air for fibre counting	Documented In-House Method GDP4 based on HSG 248	B, E, G, H*, I*, J*, K*, L*, M*, N, O, P
	Fibre counting	Documented In-House Method GDP4, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B, E, G, H*, I*, J*, K*, L*, M*, N, O, P
	4 Stage Clearance Process	Documented In-House Method GDP4, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	H*, I*, J*, K*, L*, M*, N
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method GDP1 based on HSG 248	H*, I*, J*, K*, L*, M*, N
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method GDP3 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, B, C, D, F, G, H*, I*, J*, K*, L*, M*, Q
END			