


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS TESTING 1780 Accredited to ISO/IEC 17025:2017	Fibre Clear Analytical Ltd	
	Issue No: 020 Issue Date: 01 February 2022	
	Suit 7, Sopwith House Sopwith Crescent Wickford Business Park Essex SS11 8YU	Contact: Ian Elder Tel: +44 (0)1268 571058 E-Mail: ianelder@fibreclear.org Website: www.fibreclear.org
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory location:

Location details	Activity	Location code	
Address Suit 7, Sopwith House Sopwith Crescent Wickford Business Park Essex SS11 8YU	Local contact Mr Ian Elder Tel: +44 (0)1268 571058 Email: ianelder@fibreclear.org	Health and Hygiene Head Office Asbestos – All Support Functions	C

Site activities performed away from the locations listed above:

Location details	Activity	Location code	
Client Premises	Local contact Mr Mr Ian Elder	Health and Hygiene	A
Mobile Testing Laboratories	Mr Ian Elder	Health and Hygiene	B



1780
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Fibre Clear Analytical Ltd
Issue No: 020 **Issue date:** 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	A, B, C
	Sampling of air for fibre counting	Documented In-House Method TM-02 based on HSG 248	
	Fibre Counting	Documented In-House Method TM-02, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method TM-02, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method TM-03 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	C

END