


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 Accredited to ISO/IEC 17025:2017	Common Services Agency - NHS National Services Scotland	
	Issue No: 014 Issue date: 24 September 2020	
	Gyle Square 1 South Gyle Crescent Edinburgh EH12 9EB	Contact: Mr N McLean Tel: +44 (0) 131 275 6562 Email: norman.mclean@nhs.scot Website: http://www.hfs.scot.nhs.uk/validation
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code	
Customer Premises - Any	Local contact Mr N McLean	Temperature	S
	Tel: +44 (0)131-275 6562	Pressure	S
	Email: norman.mclean@nhs.scot Website: www.hfs.scot.nhs.uk/validation	Elapsed Time	S



1957
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Common Services Agency - NHS National Services Scotland

Issue No: 014 Issue date: 24 September 2020

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
STERILIZERS AND AUTOCLAVES (excluding LTS Disinfectors, LTSF (formaldehyde) sterilizers, Ethylene oxide sterilizers, and Irradiation sterilizers)		HTM 2010:Part 3:1994 Sections as shown below:	
Clinical sterilizers, Laboratory sterilizers, and Media preparators	Annual safety checks Chamber Temperature Profile Tests Thermometric Tests for: - Performance qualification - Performance requalification Steam Quality Tests Chamber Integrity Tests Automatic Control Tests	Section 5.8 Section 7 Section 8 Section 9 Section 11 Section 12	S
Porous Load Sterilizers	Thermometric Tests Load Dryness Tests Steam Penetration Tests	Section 13	S
Fluid Sterilizers	Thermometric Tests	Section 14 (excluding paras 14-32, Coolant Quality Tests, and 14.4, Heat Exchanger Integrity Tests)	S
Sterilizers for Unwrapped Instruments and Utensils	Thermometric Tests	Section 15	S
Dry-heat Sterilizers	Thermometric Tests	Section 16 (excluding para 16.13, Air Filter Integrity Tests)	S
Laboratory Sterilizers and Culture Media Preparators	Thermometric Tests Thermal Door-lock Override Test	Section 19 (excluding para 19.78, Reheat and Dispensing Tests)	S



1957
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Common Services Agency - NHS National Services Scotland
Issue No: 014 Issue date: 24 September 2020

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
STEAM STERILIZERS Porous Load Sterilizers Laboratory Sterilizers (Autoclaves)	Validation and verification tests	SHTM 01-01: Part C: 2018 Sections as shown below: Section 3.7 (All tests listed in Table 2)	S
	Sampling of water and steam – for field and laboratory analysis	Section 4.158	S
	Physical steam quality tests	Section 4.222	S
	Validation and verification tests	SHTM 01-02: Part C: 2020 Sections as shown below: Section 3.11 (All tests listed in Table 2, with the exception of sound pressure test)	S
	Sampling of water and steam – for field and laboratory analysis	Section 4.143	S
	Physical steam quality tests	Section 4.216	S



1957
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Common Services Agency - NHS National Services Scotland
Issue No: 014 Issue date: 24 September 2020

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WASHER DISINFECTORS Surgical Instrument Washer Disinfectors	Validation and verification tests	SHTM 01-01: Part D: 2018 Sections as shown below: Section 3.23 (All tests listed in Table 4, with the exception of air filter quality test and sound pressure test) and Section 3.26 (All tests listed in Table 5)	S
	Sampling of water – for laboratory analysis	Section 7.4	S
Laboratory Washer Disinfectors	Validation and verification tests	SHTM 01-02: Part D: 2020 Sections as shown below: Section 3.24 (All tests listed in Table 4, with the exception of air filter quality test and sound pressure test) and Section 3.27 (All tests listed in Table 5)	S
	Sampling of water – for laboratory analysis	Section 6.4	S
END			