

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 1993 Accredited to ISO/IEC 17025:2017	Aerotech Laboratories Ltd	
	Issue No: 026 Issue date: 08 June 2021	
	Unit 20 Mercia Business Village Torwood Close Westwood Business Park Coventry CV4 8HX	Contact: Mr M Ricketts Tel: +44 (0)24 7647 4474 Fax: +44 (0)24 7647 4473 E-Mail: support@aerotechlabs.co.uk Website: www.aerotechlabs.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 20 Mercia Business Village Torwood Close Westwood Business Park Coventry CV4 8HX Local contact Mr M Ricketts Tel: +44 (0)24 7647 4474 Fax: +44 (0)24 7647 4473 Email: support@aerotechlabs.co.uk Website: www.aerotechlabs.co.uk	Chemical Analysis	A
Address Unit 11 Portway Close Off Torrington Avenue Tile Hill Coventry CV4 9UY Local contact Mr M Ricketts Tel: +44 (0)24 7647 4474 Fax: +44 (0)24 7647 4473 Email: support@aerotechlabs.co.uk Website: www.aerotechlabs.co.uk	Salt Spray Testing	B
Address Unit 17 Mercia Business Village Torwood Close Westwood Business Park Coventry CV4 8HX Local contact Mr M Ricketts Tel: +44 (0)24 7647 4474 Fax: +44 (0)24 7647 4473 Email: support@aerotechlabs.co.uk Website: www.aerotechlabs.co.uk	Fluorescent Penetrant Test	C



1993
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Aerotech Laboratories Ltd
Issue No: 026 Issue date: 08 June 2021

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Ferrous & non-ferrous metals Protective coatings Sodium chloride for use in salt spray testing	Neutral Salt spray	ASTM B117-19 ISO 9227:2017 4 off cabinets Largest working size:1465 mm (W) x 1040 mm (D) x 720 mm (H)	B
	Copper, nickel, iodide, lead	Stripping Voltammetry (TA216 series in-house methods)	A
	Fluoride, bromide, iodide	Ion chromatography	A
	Copper, bromide, sodium chloride, anti-caking agent	Volumetric and colorimetric analysis	A
Fluorescent penetrant consumables for NDT	Fluorescent brightness Removability (Washability) Water content Sensitivity	Documented in-house methods (TA142 Series) based on ASTM E1417/E1417M-16, Section 7.8 and referencing ASTM E1135-19 ASTM D95-13	C
	Plating solutions, anodising solutions including sulphuric & tartaric acids, etching solutions, cleaners, pre-treatment solutions and other aqueous solutions	Documented in-house methods (TA Series)	
	Metallic elements	Atomic absorption, spectrophotometry, volumetric, gravimetric and colorimetric analysis	A
	Inorganic compounds and radicals	Volumetric, gravimetric and colorimetric analysis, ion-chromatography and auto-titration	A
	Acid, alkali and other chemical concentrations	Volumetric and gravimetric, ion-chromatography and auto-titration	A
	Chloride, fluoride, sulphate and other anion content	Volumetric and gravimetric' ion-chromatography and auto-titration	A



1993
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Aerotech Laboratories Ltd
Issue No: 026 Issue date: 08 June 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Process waters, wastewaters and effluents		Documented in-house methods (TA171 Series)	A
	Fluoride, chloride sulphate	Ion chromatography	A
	Chloride	Titrimetry	A
	Aluminium, cadmium, copper, chromium, iron, lead, nickel, zinc	Flame atomic absorption spectrophotometry.	A
	Chemical oxygen demand	Spectrophotometry	A
	Free cyanide	Spectrophotometry	A
	Silica	Spectrophotometry	A
	Suspended solids	Gravimetry	A
	Conductivity and total dissolved solids	Conductivity meter	A
	pH	pH meter	A
Oil samples	Water content	Documented in-house method (TA188-1) using Karl Fischer determination technique	A

END