


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 2054 Accredited to ISO/IEC 17025:2017	Airborne Environmental Consultants Limited	
	Issue No: 020 Issue date: 14 January 2020	
	23 & 27 Wheelforge Way Ashburton Point Trafford Park Manchester M17 1EH	Contact: Mr R Harris Tel: +44 (0)161 872 7111 Fax: +44 (0)161 872 7112 E-Mail: sales@aec.uk.net Website: www.aec.uk.net

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 23 & 27 Wheelforge Way Ashburton Point Trafford Park Manchester M17 1EH	Local contact Mr R Harris	Health and Hygiene A
Address Fernwood House The Hilway Mounthessing Brentwood CM15 0UE	Local contact Mr R Harris Tel: 0203 384 6175 Fax: 0203 468 6904	Health and Hygiene C

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	D



2054
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airborne Environmental Consultants Limited
Issue No: 020 Issue date: 14 January 2020

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG248: February 2005 Membrane Filter Method using Phase Contrast Microscopy (PCM)	A,B, D
	Sampling of air for fibre counting	HSG248: February 2005 by Documented In-House Method AEC 1	B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG248: February 2005 (Documented In-House Procedure)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure AEC 2 using stereo polarised light - microscopy, microscopy and dispersion staining	A, C
Asbestos in soils – Sediment and aggregate (fibre screening and identification)	Sampling of bulk materials for asbestos identification	Documented In-House Method AEC 2	B
	Abestos Fibre Screening and identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method AEC 2 using stereo-microscopy, polarised light optical microscopy and dispersion staining.	A



2054
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airborne Environmental Consultants Limited
Issue No: 020 Issue date: 14 January 2020

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Asbestos in soils – Sediment and aggregate (screening/identification/quantification)	Abestos Fibre Screening/identification and quantification of asbestos content of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method AEC 2 using stereo-microscopy, polarised light optical microscopy and dispersion staining.	A
END			