

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |   |   |
|--|---|---|
| <br><b>2146</b><br><br>Accredited to<br><b>ISO/IEC 17025:2017</b> | <b>Ethos Environmental Ltd</b>  |   |
|  | Issue No: 015    Issue date: 29 September 2021  |   |
|  | <b>Unit 16</b><br>Dumbryden Industrial Estate<br>32 Dumbryden Road<br>Edinburgh<br>EH14 2AB | <b>Contact: Dr B Gardner</b><br>Tel: +44 (0)131 453 5111<br>Fax: +44 (0)131 453 6111<br>E-Mail: <a href="mailto:info@ask-ethos.co.uk">info@ask-ethos.co.uk</a><br>Website: <a href="http://www.ask-ethos.co.uk">www.ask-ethos.co.uk</a> |

Testing performed by the Organisation at the locations specified below

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details   | Activity                             | Location code               |
|--|--------------------------------------|-----------------------------|
| <b>Address</b><br>Unit 16<br>Dumbryden Industrial Estate<br>32 Dumbryden Road<br>Edinburgh<br>EH14 2AB | <b>Local contact</b><br>Dr B Gardner | Health and Hygiene<br><br>A |

#### Site activities performed away from the locations listed above:

| Location details                                       | Activity   | Location code |
|--|--|---------------|
| Premises including domestic, commercial and industrial | Health and Hygiene   | B             |
| Any suitable customer premises                         | Post completion sound insulation testing for the Building (Scotland) Regulations 2004 using the guidance given in Technical Handbook – Domestic, Section 5, Noise for Domestic Buildings – 2016, 2017 & 2019 | C             |



2146  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Ethos Environmental Ltd**  
**Issue No: 015 Issue date: 29 September 2021**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Materials/Products tested  | Type of test/Properties measured/Range of measurement  | Standard specifications/ Equipment/Techniques used  | Location Code |
|--|--|---|---------------|
| ASBESTOS FIBRES IN AIR   | <u>Health and Hygiene</u>  | Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)                   |               |
|  | Fibre Counting   | Documented In-House Procedure (26) based on Membrane Filter Method using Phase Contrast Microscopy (PCM)                              | A, B          |
|  | Sampling of air for fibre counting   | HSG 248:February 2005 (Documented In-House Procedure 20) based on Membrane Filter Method using Phase Contrast Microscopy (PCM)        | B             |
| ASBESTOS IN BULK MATERIALS Including materials and products suspected of containing asbestos | 4 Stage Clearance Process  | HSG 248:February 2005 (Documented In-House Procedure 20) based on Membrane Filter Method using Phase Contrast Microscopy (PCM)        | B             |
|  | Identification of:<br>Amosite<br>Chrysotile<br>Crocidolite<br>Fibrous Actinolite<br>Fibrous Anthophyllite<br>Fibrous Tremolite | HSG 248:February 2005 by Documented In-House Procedure 27 using stereo-microscopy, polarised light microscopy and dispersion staining | A             |
|  | Sampling of bulk materials for asbestos identification   | Documented In-House Procedure 47  | B             |



2146  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Ethos Environmental Ltd**  
**Issue No: 015 Issue date: 29 September 2021**

Testing performed by the Organisation at the locations specified

| Materials/Products tested                                  | Type of test/Properties measured/Range of measurement  | Standard specifications/ Equipment/Techniques used   | Location Code |
|--|--|--|---------------|
| <b>Building Elements such as partitions, walls, floors</b> | <u>Acoustics</u><br>Field measurement of airborne sound insulation, ie Level Difference and related factors              | Post completion sound insulation testing for the Building (Scotland) Regulations 2004 using the guidance given in Technical Handbook – Domestic, Section 5, Noise for Domestic Buildings – 2016, 2017 & 2019 | C             |
| <b>Building Elements - floors</b>                          | Field measurement of impact sound insulation of floors between rooms, ie Impact Sound Pressure Level and related factors | Post completion sound insulation testing for the Building (Scotland) Regulations 2004 using the guidance given in Technical Handbook – Domestic, Section 5, Noise for Domestic Buildings – 2016, 2017 & 2019 | C             |
| END  |  |  |               |