


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>2230</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>Asbestos Analytical Services</b>	
	Issue No: 013    Issue date: 12 October 2016	
	<b>Unit 1</b> Shawfarm Road Prestwick KA9 2TR	<b>Contact: Mr D Espie</b> Tel: +44 (0)1292 475176 Fax: +44 (0)1292 474845 E-Mail: info@as-ayr.co.uk Website: aasayr.co.uk
Testing performed by the Organisation at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Unit 1 Shawfarm Road Prestwick KA9 2TR	<b>Local contact</b> Mr D Espie	Health and Hygiene  A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B



2230

Accredited to  
ISO/IEC 17025:2005

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Asbestos Analytical Services

Issue No: 013 Issue date: 12 October 2016

Testing performed by the Organisation at the locations specified

#### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	Membrane Filter Method using Phase Contrast Microscopy (PCM) based upon HSG 248	A, B
	Sampling of air for fibre counting	Documented In-House Method No LSD.ASB.02 based upon HSG 248	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Membrane Filter Method using Phase Contrast Microscopy (PCM) based upon HSG 248	B
	<u>Health and Hygiene</u>		
	Sampling of bulk materials for asbestos identification	Documented In-House Method No LSD.ASB.03 based on HSE HSG 264	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method No LSD.ASB.01 using stereo-microscopy, polarised light microscopy and dispersion staining based upon HSG 248	A
END			