


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>2418</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>Acorn Analytical Services Ltd</b>  <b>Issue No: 019 Issue date: 21 May 2025</b>	
	<b>The Old Print Works</b> <b>Carr Street</b> <b>Cleckheaton</b> <b>West Yorkshire</b> <b>BD19 5HG</b>	<b>Contact: James Slinger</b> <b>Tel: +44 (0) 1924 443 552</b> <b>E-Mail: James@acorn-as.com</b> <b>Website: www.acorn-as.com</b>
Testing performed by the Organisation at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<b>Address</b> The Old Print Works Carr Street Cleckheaton West Yorkshire BD19 5HG	<b>Local contact</b> Andrew Jackson  Tel: +44 (0) 1924 443 552 E-Mail: andrew.jackson@acorn-as.com Website: www.acorn-as.com	Health and Hygiene Head Office  Asbestos – All Support Functions	A
<b>Address</b> 25 Charter Gate Quarry Park Close Moulton Park Industrial Estate Northampton NN3 6QB	<b>Local contact</b> Paul Knights  Tel: +44 (0) 1604 648928 E-Mail: Paul.knights@acornasbestos.co.uk Website: www.acorn-as.com	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none"><li>Quality Audit</li><li>Administration</li><li>Contract Review</li><li>Scheduling</li><li>Personnel</li><li>Equipment</li><li>Measurement</li><li>Traceability</li><li>Reporting</li></ul>	C

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises (Cleckheaton)	Health and Hygiene	B
Client Premises (Northampton)	Health and Hygiene	D
Mobile Laboratory Vans	Health and Hygiene	E



2418  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Acorn Analytical Services Ltd**  
**Issue No: 019 Issue date: 21 May 2025**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method Quality Manual – Appendix I based on HSG 248	A, B, C, D, E
	Fibre counting	Documented In-House Method Quality Manual – Appendix I, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C, D, E
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method Quality Manual – Appendix 1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B, D
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method Quality Manual – Appendix 3 based on HSG 248	B, D
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method Quality Manual – Appendix 2 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, C
END			