

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 Accredited to ISO/IEC 17025:2017	Enquin Environmental Limited	
	Issue No: 016 Issue date: 20 June 2024	
	1 Ringside Business Centre Heol Y Rhosog Wentloog Cardiff CF3 2EW	Contact: Amie Charlton Tel: +44 (0)29 2053 5090 E-Mail: acharlton@enquin.co.uk Website: www.enquin.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address 1 Ringside Business Centre Heol Y Rhosog Wentloog Cardiff CF3 2EW	Local contact Amie Charlton Tel: +44 (0)29 2053 5090 Email: asbestos@enquin.co.uk Website: www.enquin.co.uk	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



2604
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Enquin Environmental Limited
Issue No: 016 **Issue date:** 20 June 2024

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method in the Procedures Manual, TP02 (Section 2) based on HSG 248	A, B
	Fibre counting	Documented In-House Method in the Procedures Manual, TP02 (Section 2), Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method in the Procedures Manual, TP02 (Section 2), Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method in the Procedures Manual, TP04 (Section 4) based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method in the Procedures Manual, TP03 (Section 3) using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			