

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



2671

Accredited to
ISO/IEC 17025:2017

Fourstage Environmental Ltd

Issue No: 014

Issue date: 01 February 2022

5a Clapgun Street
Castle Donington
Derbyshire
DE74 2LE

Contact: Mr D Higbed
Tel: +44 (0)1332 812549
Fax: +44 (0)1332 814908
E-Mail: info@fourstage.co.uk
Website: www.fourstage.co.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 5a Clapgun Street Castle Donington Derbyshire DE74 2LE	Local contact Mr D Higbed Tel: +44 (0)1332 812549 Fax: +44 (0)1332 814908 Email: info@fourstage.co.uk	Health and Hygiene Head Office Asbestos – All Support Functions A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene Asbestos	B
Mobile Testing Laboratories	Health and Hygiene	C



2671
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Fourstage Environmental Ltd
Issue No: 014 Issue date: 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method DOC/002 based on HSG 248	A, B
	Fibre counting	Documented In-House Method DOC/002, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
	4 Stage Clearance Process	Documented In-House Method DOC/002, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method DOC/101 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A

END