

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



2705

Accredited to
ISO/IEC 17025:2017

Quality Consultants (NI) Ltd

Issue No: 018 Issue date: 26 March 2025

2 Cyprus Gardens
Belfast
BT5 6FB

Contact: Mr I Harper
Tel: +44 (0)28 9048 4905
E-Mail: iharper@qcni.co.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 2 Cyprus Gardens Belfast BT5 6FB Local contact Mr I Harper Tel: +44 (0) 28 9048 4905 Email: iharper@qcni.co.uk	Health and Hygiene Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



2705
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Quality Consultants (NI) Ltd
Issue No: 018 Issue date: 26 March 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method SOP03 based on HSG 248	A, B
	Fibre counting	Documented In-House Method SOP03 Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
	4 Stage Clearance Process	Documented In-House Method, SOP03 Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method, SOP01 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			