

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING</p> <p>2719</p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<p><b>Environmental Essentials Ltd</b></p> <p>Issue No: 034      Issue date: 27 June 2022</p>	
	<p><b>Unit 3, Arlington Court</b> <b>Silverdale Enterprise Park</b> <b>Cannel Row</b> <b>Newcastle-under-Lyme</b> <b>ST5 6SS</b></p>	<p><b>Contact: R Rushton</b> <b>Tel: +44 (0)845 456 9953</b> <b>Fax: +44 (0)845 456 9954</b> <b>E-Mail: rrushton@environmentalessentials.co.uk</b> <b>Website: www.environmentalessentials.co.uk</b></p>
<p><b>Testing performed by the Organisation at the locations specified below</b></p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b> Head Office</p> <p>Environmental Essentials Ltd Unit 3 Arlington Court Silverdale Enterprise Park Cannel Row Newcastle-under-Lyme ST5 6SS</p> <p><b>Local contact</b> Danny Kirbyshire Tel: +44 (0)845 456 9953 Fax: +44 (0)845 456 9954 Email: dkirbyshire@environmentalessentials.co.uk Website: www.environmentalessentials.co.uk</p>	<p>Health and Hygiene Head Office</p> <p>Asbestos – All Support Functions</p>	A
<p><b>Address</b> Scotland Office</p> <p>Environmental Essentials Ltd Suite 1A International House Hamilton International Technology Park Stanley Boulevard High Blantyre G72 0BN</p> <p><b>Local contacts</b> Angela McPhillips Tel: +44 (0)169 840 4530 Fax: +44 (0)845 456 9954 Email: amcphillips@environmentalessentials.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• Quality Audit</li> <li>• Administration</li> <li>• Contract Review</li> <li>• Scheduling</li> <li>• Personnel</li> <li>• Equipment</li> <li>• Measurement</li> <li>• Traceability</li> <li>• Reporting</li> </ul>	I
<p><b>Address</b> South East Office</p> <p>Environmental Essentials Ltd First Floor ROC House Faraday Court Faraday Road Crawley West Sussex RH10 9PU</p> <p><b>Local Contacts</b> Katie French Tel: +44 (0)7796 715932 Fax: +44 (0)845 456 9954 Email: kfrench@environmentalessentials.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• Quality Audit</li> <li>• Administration</li> <li>• Contract Review</li> <li>• Scheduling</li> <li>• Personnel</li> <li>• Equipment</li> <li>• Measurement</li> <li>• Traceability</li> <li>• Reporting</li> </ul>	F



2719  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Environmental Essentials Ltd**  
**Issue No: 034 Issue date: 27 June 2022**

Testing performed by the Organisation at the locations specified

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	A, I, F
Mobile Testing Laboratories	Health and Hygiene	A, I, F



2719  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Environmental Essentials Ltd**  
**Issue No: 034 Issue date: 27 June 2022**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method based on HSG 248	A, I, F
	Fibre counting	Documented In-House Method, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, I, F
	4 Stage Clearance Process	Documented In-House Method, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, I, F
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method based on HSG 248	A, I, F
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, I, F

END