


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 2791 Accredited to ISO/IEC 17025:2017	BDA Surveying Ltd	
	Issue No: 011 Issue date: 01 February 2022	
	Chapel Street Oadby Leicester LE2 5AD	Contact: Mr R Hubbard Tel: +44 (0)116 271 9179 Fax: +44 (0)116 271 6478 E-Mail: info@bdasurveying.co.uk Website: www.bdasurveying.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Chapel Street Oadby Leicester LE2 5AD Local contact Mr R Hubbard Tel: +44 (0)116 271 9179 Fax: +44 (0)116 271 6478 Email: info@bdasurveying.co.uk Website: www.bdasurveying.co.uk	Asbestos Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Asbestos	C
Mobile Testing Laboratories	Asbestos	D



2791
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

BDA Surveying Ltd
Issue No: 011 Issue date: 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method based on HSG 248	C
	Fibre counting	Documented In-House Method, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, C, D
ASBESTOS IN BULK MATERIALS Including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	C,
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method based on HSG 248	A, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A,
END			