

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS TESTING 2798</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Bellingham & Stanley Limited</h3> <p>Issue No: 006 Issue date: 23 December 2019</p>	
	<p>Longfield Road Tunbridge Wells Kent TN2 3EY</p>	<p>Contact: Mr Andrew Darkins Tel: +44 (0)1892 500400 Fax: +44 (0)1892 543115 E-Mail: sales.bs.uk@xyleminc.com Website: www.bellinghamandstanley.com</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
LIQUIDS Food products and ingredients, chemical or petroleum based products	Refractive Index (1.3 - 1.62 at 20.0°C and 589.3 nm)	Documented In-House Methods QL-105
	°Brix (0 - 60 at 20.0°C and 589.3 nm)	QL-104
GLASSES/PLASTICS	Refractive Index (1.3 - 1.62 at 20.0°C and 589.3 nm)	QL-114
LIQUIDS/SOLUTIONS/QUARTZ PLATES	Optical rotation on °A or °Z scales at 589.44 nm and 10 - 30°C -11.77 - 36.36°A -34 - 105°Z	QL-107 QL-110
<p>END</p>		