


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS TESTING</p> <p>4013</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Birds Eye Limited</p> <p>Issue No: 014 Issue date: 28 August 2019</p>	
	<p>Whapload Road Lowestoft Suffolk NR32 1XQ</p>	<p>Contact: Caroline Hytch Tel: +44 (0)1502 504998 Fax: +44 (0)1502 504516 E-Mail: caroline.hytch@birdseye.co.uk</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
General Foods, including Frozen and Chilled Foods and Materials, Meats, Fish, Vegetables, Rice and Pasta	<p><u>Microbiological Tests</u></p> <p>Enumeration of:</p> <p>Aerobic colony count at 30°C</p> <p><i>Escherichia coli</i> (βeta-glucuronidase-positive)</p> <p>Coagulase-positive staphylococci, including <i>Staphylococcus aureus</i></p> <p>Presumptive <i>Bacillus cereus</i></p> <p><i>Listeria monocytogenes</i></p> <p><i>Salmonella</i> spp</p>	<p>Documented In-House Methods:</p> <p>MCM-010 by Petrifilm MCM-009 by PCA pour plate incubated at 30°C for 48h</p> <p>MCM-005 by Petrifilm: SEC</p> <p>MCM-021 based on BS EN ISO 6888-1:1999+A2:2018 using confirmation by rabbit plasma coagulation</p> <p>MCM-020 in house method by spread plate using chromogenic agar incubated at 30°C for 18-24h</p> <p>MCM-004 based on BS EN ISO 11290-2:2017, confirmation using BioMerieux RAPIDEC Lmono</p> <p>MCM-008. Detection by VIDAS and confirmation of positive samples by: Selective plating onto XLD and other chromogenic agar Serological confirmation of colonies with Salmonella Latex Kit and Biochemical confirmation by RapID ONE</p>
<p>END</p>		