


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>4097</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>ASI Environmental Limited</b>	
	Issue No: 011	Issue date: 01 February 2022
	6 Testwood Lane Totton Southampton SO40 3AP	Contact: Mr D Grocott Tel: +44 (0)2380 427750 Fax: +44 (0)2380 668755 E-Mail: david@asi.co.com or info@asi.co.com Website: www.asi-asbestos-surveys.co.uk
Testing performed by the Organisation at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> 6 Testwood Lane Totton Southampton SO40 3AQ  <b>Local contact</b> Mr D Grocott  Tel: +44 (0)2380 427750 Fax: +44 (0)2380 688755 Email: info@asi-surveys.co.uk	Health and Hygiene Head Office  Asbestos – All Support Functions	A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



4097  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**ASI Environmental Limited**  
Issue No: 011 Issue date: 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method Procedure SOP 02 based on HSG 248	A, B, C
	Fibre counting	Documented In-House Method SOP 03, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method SOP 04, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method TP02 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method TP01 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			