


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>4211</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>NHS Greater Glasgow & Clyde, Microbiology Department</h3> <p>Issue No: 009 Issue date: 26 February 2021</p>	
	<p>Clinical Microbiology Department Level 4 New Lister Building Glasgow Royal Infirmary 10-16 Alexandra Parade Glasgow G31 2ER</p>	<p>Contact: Mr David Jordan Tel: +44 (0)141 2018570 Fax: +44 (0)141 2018535 E-Mail: David.Jordan@ggc.scot.nhs.uk</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATER (Potable, natural and process waters)	<p><u>Microbiology Testing</u></p> <p><i>Legionella</i> spp - Detection and enumeration</p> <p>Identification of: <i>Legionella pneumophila</i> serogroup 1 <i>L. pneumophila</i> serogroups 2-14 <i>Legionella</i> species</p>	<p>Documented In-House method LP532 based on ISO 11731:2017</p>
Potable, recreation waters (man made) and renal dialysis waters	<p>Total Viable Count</p> <p>Coliforms & <i>Escherichia coli</i></p> <p><i>Pseudomonas</i> species and <i>Pseudomonas aeruginosa</i></p>	<p>Documented In-House method LP538 based on the Microbiology of Drinking Water series, and the Renal Association Guidelines and ISO 13959:2015</p> <p>Documented In-House method LP538 based on ISO 9308-01:2014 (E), and the Renal Association Guidelines and ISO 13959:2015</p> <p>Documented In-House method LP538 based on the Microbiology of Drinking Water series, and the Renal Association Guidelines and ISO 13959:2015</p>
Endoscopy rinse waters	<p>Total Viable Count</p> <p><i>Pseudomonas aeruginosa</i></p>	<p>Documented In-House method LP538 based on the the requirements of HTM -01-06</p> <p>Documented In-House method LP538 based on the the requirements of HTM -01-06</p>



4211
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Microbiology Department Glasgow Royal Infirmary
Issue No: 009 Issue date: 26 February 2021

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATER (Endoscopy rinse waters) (cont'd)	<u>Microbiology Testing</u> (cont'd) Atypical Mycobacteria	Documented In-House method LP538 based on the requirements of HTM -01-06
MICRO ORGANISMS IN AIR Air samples	Enumeration Total Viable Count Yeasts and Moulds Total Viable Count Yeasts and Moulds	Settle plates, contact plates and air sampling using documented In-House method LP539 Particle counting using documented In-House method LP539
END		