


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS TESTING 4234 Accredited to ISO/IEC 17025:2017	Asbestos Solution Providers Ltd	
	Issue No: 009 Issue date: 01 February 2022	
	906 Yeovil Road Slough Trading Estate Slough Berkshire SL1 4JG	Contact: Mr J Johal Tel: +44 (0)1753 694 422 E-Mail: enquiries@aspltd.biz Website: www.aspltd.biz
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 906 Yeovil Road Slough Berkshire SL1 4JG Local contact Mr J Johal Tel: +44 (0) 1753 694 422 E-Mail: enquiries@aspltd.biz Website: www.aspltd.biz	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



4234
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Asbestos Solution Providers Ltd
Issue No: 009 Issue date: 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method Technical Guidance Note 1 based on HSG 248	A, B
	Fibre counting	Documented In-House Method Technical Guidance Note 1 using Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method Technical Guidance Note 1 using Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method Technical Guidance Note 2 based on HSG 248	B
END			