

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 4243 Accredited to ISO/IEC 17025:2017	Occupational Management Consultants Ltd	
	Issue No: 006 Issue date: 01 February 2022	
	Unit 8 Hoyer Industrial Park 517 Leeds Road Huddersfield HD2 1YJ	Contact: Debbie McGregor-Cheers Tel: +44 (0)1484 425043 Fax: +44 (0)1484 425044 E-Mail: debbie@om-consultants.co.uk Website: https://www.om-consultants.co.uk/
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code	
Address Unit 8 Hoyer Industrial Park 517 Leeds Road Huddersfield HD2 1YJ	Local contact Debbie McGregor-Cheers	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



4243
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Occupational Management Consultants Ltd
Issue No: 006 **Issue date:** 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method PRO 024 - Air Monitoring Procedure based on HSG 248	B
	Fibre counting	Documented In-House Method PRO 024 – Air Monitoring Procedure, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method PRO 025 – 4 Stage Clearance Procedure, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for asbestos identification	Documented In-House Method PRO 033 – Bulk Sampling Procedure based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method PRO 058 – Bulk Analysis Procedure using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			