## **Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service** 

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



## DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
Refinery Gas Analysis	Composition % mole and apparent molecular weight	Flexible scope enabling new versions of existing accredited standard test methods to be introduced in accordance with documented in house procedure OGC-GEN-1QP-018
	Acetylene 0.01 to 0.48 Argon / Oxygen 0.01 to 5.06 iso-Butane 0.01 to 5.01 1,3-Butadiene 0.01 to 1.01 1-Butene 0.01 to 5.01 C5+ 0.01 to 1.99 Cis-2-Butene 0.01 to 1.51 Iso-Butylene 0.01 to 2.02 Carbon dioxide 0.01 to 5.02 Carbon monoxide 0.01 to 10.09 Ethane 0.01 to 24.89 Ethylene 0.01 to 19.97 Hydrogen 0.01 to 50.39 Hydrogen sulphide 0.01 to 5.03 Methane 0.01 to 39.97 iso-Pentane 0.01 to 0.99 n-Pentane 0.01 to 0.99 Propadiene 0.01 to 0.5 Propane 0.01 to 10.01 Propylene 0.01 to 10.01 Propylene 0.01 to 3.06 Carbon content Calorific value	BS EN 15984

