


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|---|---|
|  4346 Accredited to ISO/IEC 17025:2017 | RWE Generation UK plc | |
| | Issue No: 008 Issue date: 11 December 2019 | |
| | Didcot B Power Station Didcot Oxfordshire OX11 7YU | This laboratory does not accept commercial or external enquiries Website: www.rwegeneration.com |
| Testing performed by the Organisation at the locations specified below | | |

Locations covered by the organisation and their relevant activities

Laboratory locations:

| Location details | Activity | Location code | |
|---|---|---|---|
| Address Didcot B Power Station Didcot Oxfordshire OX11 7YU | Local contact This laboratory does not accept commercial or external enquiries Website: www.rwenpower.com | Support Functions: Quality System Quality Audit Administration Sampling and Testing: Chemical Tests | A |

Site activities performed away from the locations listed above:

| Location details | Activity | Location code |
|---|------------------------|---------------|
| Consented water discharges for RWE NPower PLC, Didcot | Environmental Sampling | B |



4346
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RWE Generation UK plc
Issue No: 008 Issue date: 11 December 2019

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used | Location Code |
|---|--|---|---------------|
| WATERS - River water and site industrial effluent | <u>Sampling for chemical analysis conducted in-house</u> | Documented In-House Methods | B |
| | <u>Chemical Analysis</u> | | |
| | Total Chlorine (0 - 5 mg/l) | Documented In-House Method to BS EN ISO 7393-2:2000 (DID-LAB-008) | A |
| | Oil in Water (0 - 200 mg/l) | Documented In-House Method using Partition Infrared based on standard method 5520C for examination of water and waste water (DID-LAB-009) | A |
| | pH (0 - 14) | Documented In-House Method to BS 6068-2.5:1995 (DID-LAB-006) | A |
| | Suspended solids (≥2 mg/l) | Documented In-House Method to BS EN 872:2005 (DID-LAB-010) | A |
| | Chloride Sulphate | Documented In-House Method to BS EN ISO 10304-1:1995 (DID-LAB-007) | A |
| END | | | |