


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 4376 Accredited to ISO/IEC 17025:2017	Airsafe Analytical Limited	
	Issue No: 007 Issue date: 03 February 2022	
	14 Normandy Street Alton Hampshire GU34 1BX	Contact: Mr Richard Wren Tel: +44 (0) 1420 88883 E-Mail: richard@airsafe.org.uk Website: www.airsafe.org.uk
Testing performed at the above address only		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<u>Health and Hygiene</u> Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021 Documented In-House Method Documented in 'Airsafe Analytical Procedures Manual,' using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248
END		