

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>4414</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Cobb Europe Ltd</h3> <p>Issue No: 011 Issue date: 12 December 2018</p>	
	<p>Quality Control Centre Greens Road Dereham NR20 3TG</p>	<p>Contact: Mrs G Walker Tel: +44 (0)1362 708514 E-Mail: gemma.walker@cobb-europe.com</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>ANIMAL FEEDS, ENVIRONMENTAL SAMPLES/MATERIALS AND FAECAL SAMPLES AS SPECIFIED</p> <p>Chick Box Liners, Boot Swabs, Composite Faeces, Dust, Dust Swabs, Meconium (optional)</p>	<p><u>Microbiological Tests</u></p> <p>Detection of <i>Salmonella</i> spp (motile)</p>	<p>Documented In-House Procedures</p> <p>QCP.07.000, QCP.07.001, QCP.07.004 documented in house method based on BS EN ISO 6579- 1:2017 (employing MSRV) for matrices as required for the purposes of testing in accordance with Control of Salmonella in Poultry Orders*: CSPO 2007 (2008 NI) CSBO 2009 (2009 NI) CSTO 2009 (2010 NI) Confirmation using group level serology and biochemical reaction (oxidase and Remel RapID™ ONE) and Remel agglutination antisera reagents * refer to www.gov.uk for current laboratory approval status</p>
<p>Fluff, Feed/Raw Materials, Litter</p>	<p>Detection of <i>Salmonella</i> spp (motile)</p>	<p>QCP.07.000, QCP.07.001, QCP.07.004 documented in house method based on BS EN ISO 6579-1:2017, confirmation using group level serology and biochemical reaction (oxidase and Remel RapID™ ONE) and Remel agglutination antisera reagents</p>



4414
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Cobb Europe Ltd

Issue No: 011 **Issue date:** 12 December 2018

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
POULTRY BLOOD SAMPLES	<u>Serological Tests</u> Detection of antibodies to: <i>Mycoplasma gallisepticum</i> <i>Salmonella pullorum</i> <i>Mycoplasma synoviae</i> Avian influenza viruses <u>Molecular Biology Tests</u>	QCP.04.001 Rapid Slide Agglutination Test in accordance with DEFRA Poultry Health Scheme Regulation 5, Part 1 and Animals and Animal Products Regulations 2007 SI2007/1621 QCP.04.001 Rapid Slide Agglutination Test in accordance with DEFRA Poultry Health Scheme Regulation 5, Part 1 and Animals and Animal Products Regulations 2007 SI2007/1621 QCP.04.001 Rapid Slide Agglutination Test QCP.03.000 using IDEXX ELISA kit
MSRV bacterial isolates	Detection of <i>Salmonella</i> spp DNA	QCP.02.004 and supporting procedures using manual extraction and real time PCR using BioRad thermal cyclers
Cleft Palate Swabs	Detection of <i>Mycoplasma synoviae</i> DNA Detection of <i>Mycoplasma gallisepticum</i> DNA Detection of Infectious Laryngotracheitis DNA	QCP.02.004 and supporting procedures using manual extraction and real time PCR using BioRad thermal cyclers QCP.02.004 and supporting procedures using manual extraction and real time PCR using BioRad thermal cyclers QCP.02.004 and supporting procedures using manual extraction and real time PCR using BioRad thermal cyclers
END		