

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



Accredited to
ISO/IEC 17025:2017

Nichol Associates Limited

Issue No: 007

Issue date: 01 February 2022

Unit 19

Blue Sky Way

Monkton Business Park South

Hebburn

Tyne and Wear

NE31 2EQ

Contact: Mr D Nichol

Tel: +44 (0)191 438 5432

Fax: +44 (0)191 469 9562

E-Mail: davenichol@nicholassociates.co.uk

Website: www.nicholassociates.co.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 19 Blue Sky Way Monkton Business Park South Hebburn Tyne and Wear NE31 3EQ Local contact Mr D Nichol Tel: +44 (0)191 438 5432 Fax: +44 (0)191 469 9562 Email: davenichol@nicholassociates.co.uk Website: www.nicholassociates.co.uk	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



4502
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Nichol Associates Limited
Issue No: 007 Issue date: 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method Operating Procedure OP03 based on HSG 248	A, B
	Fibre counting	Documented In-House Method Operating Procedure OP03, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
	4 Stage Clearance Process	Documented In-House Method Operating Procedure OP03, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method Operating Procedure OP02 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			