

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>4525</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3>Slicker Recycling Limited</h3> <p><b>Issue No: 016    Issue date: 22 October 2018</b></p>	
	<p><b>Barracks Road</b> <b>Sandy Lane Industrial Estate</b> <b>Stourport-on-Severn</b> <b>Worcestershire</b> <b>DY13 9RW</b></p>	<p><b>Contact: Mr Amir Sharif</b> <b>Tel: +44 (0) 1473 256192</b> <b>Fax: +44 (0) 1473 226460</b> <b>E-Mail: amir.sharif@slickerrecycling.com</b> <b>Website: www.slickerrecycling.com</b></p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
Processed Fuel Oil	<p><u>Chemical Tests</u></p> <p>Carbon residue</p> <p>Chlorine content</p> <p>Flash Point</p> <p>Determination of Metals: Pb, Ni, Cr, Cu, Zn, As, Cd, Tl, Sb, Co, Mn, V, Hg</p>	<p>IP 13/94(14) BS 2000-13:94 ISO 6615:93</p> <p>IP 503/04 BS 2000-503:04 ISO 15597:01 Using in-house method LMP-SIP503 and WDXRF</p> <p>IP 523/15 BS 2000-523:15 BS EN ISO 3679:15 Rapid equilibrium closed cup</p> <p>IP 593/11 (modified for Hg) using in-house method LMP-SIP 593 and WDXRF</p>
Processed Fuel Oil (cont'd)	<p><u>Chemical Tests (cont'd)</u></p> <p>Total sediment</p> <p>Strong Acid number</p> <p>Sulphated ash</p>	<p>IP 375/11 BS 2000-375:11 ISO 10307-1:09</p> <p>IP 139/98(2017) BS 2000-139:98 ISO 6618:96</p> <p>IP 550/08</p>



Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Slicker Recycling Limited**

**Issue No: 16 Issue date: 22 October 2018**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
	Sulphur content	IP 336/04(2014) BS 2000-336:03 BS EN ISO 8754:03 Using in-house method LMP-SIP 336 and EDXRF
	Kinematic viscosity at 100°C	IP 71 Section 1/97 BS 2000-71:96 sect 1 BS EN ISO 3104:96
	Water content	IP 74/2000(14) BS 2000-74:00 ISO 3733:99
	Manual Sampling	Documented In-House Method MP7-9 based on IP 475/05 BS 2000-475:04 BS EN ISO 3170:04
END		