


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS TESTING 4665</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p>Defence Science and Technology Laboratory</p> <p>Issue No: 012 Issue date: 13 August 2019</p>	
	<p>Woodcock Hill Sandridge St. Albans AL4 9HQ United Kingdom</p>	<p>Contact: Sarah Newstead E-Mail: snewstead@dstl.gov.uk Website: www.dstl.gov.uk</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PRELIMINARY DRUG TESTING DEVICES	<p>Performance Tests using synthetic oral matrix for the presence of :</p> <p>Cannabinoids Group :</p> <ul style="list-style-type: none"> - Delta-9-THC <p>Benzodiazepines Group :</p> <ul style="list-style-type: none"> - Nordiazepam - Oxazepam - Temazepam - Diazepam <p>Cocaine Group :</p> <ul style="list-style-type: none"> - Cocaine - Benzoylecgonine <p>Amphetamine Group :</p> <ul style="list-style-type: none"> - D-amphetamine <p>Methamphetamine Group :</p> <ul style="list-style-type: none"> - Methylamphetamine - MDA - MDMA <p>Methadone Group :</p> <ul style="list-style-type: none"> - Methadone <p>Opiates Group</p> <ul style="list-style-type: none"> - Morphine <p>Evaluation of general functionality testing relating to devices for Type-Approval</p>	<p>In house method T303 based upon Preliminary Drug Testing Devices "A Guide to Type-Approval Procedures for Preliminary Drug Testing Devices Used for Transport Law Enforcement in Great Britain" excluding testing relating to Annex B</p>



4665

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Defence Science and Technology Laboratory

Issue No: 012 Issue date: 13 August 2019

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PRELIMINARY DRUG TESTING DEVICES (cont'd)	<p>Re-assurance tests of the on-going performance of disposable components used in Type Approved Preliminary Drug Testing Devices using synthetic oral matrix for the presence of:</p> <p>Cannabinoids Group :</p> <ul style="list-style-type: none"> - Delta-9-THC <p>Cocaine Group :</p> <ul style="list-style-type: none"> - Cocaine 	<p>In house method T306 based upon Mobile Preliminary Drug Testing Devices "A Guide to Type-Approval Procedures for Mobile Preliminary Drug Testing Devices Used for Transport Law Enforcement in Great Britain" excluding testing relating to Annex B</p>
MOBILE PRELIMINARY DRUG TESTING DEVICES	<p>Performance Tests using synthetic oral matrix for the presence of :</p> <p>Cannabinoids Group :</p> <ul style="list-style-type: none"> - Delta-9-THC <p>Cocaine Group :</p> <ul style="list-style-type: none"> - Cocaine - Benzoylecgonine <p>Evaluation of general functionality testing relating to devices for Type-Approval.</p> <p>Re-assurance tests of the on-going performance of disposable components used in Mobile Type Approved Preliminary Drug Testing Devices using synthetic oral matrix for the presence of:</p> <p>Cannabinoids Group :</p> <ul style="list-style-type: none"> - Delta-9-THC <p>Cocaine Group :</p> <ul style="list-style-type: none"> - Cocaine 	<p>In house method T305 based upon Mobile Preliminary Drug Testing Devices "A Guide to Type-Approval Procedures for Mobile Preliminary Drug Testing Devices Used for Transport Law Enforcement in Great Britain" excluding testing relating to Annex B</p> <p>In house method T306 based upon Mobile Preliminary Drug Testing Devices "A Guide to Type-Approval Procedures for Mobile Preliminary Drug Testing Devices Used for Transport Law Enforcement in Great Britain" excluding testing relating to Annex B</p>



4665

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Defence Science and Technology Laboratory

Issue No: 012 Issue date: 13 August 2019

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
BREATH ALCOHOL SCREENING DEVICES (electronic Screening Devices)	Performance Tests - Moist ethanol vapour Evaluation of general functionality testing relating to devices for Type-Approval. Performance Tests - <i>In vivo</i> testing	In-house method T402 based on the Home Office "Guide to Type Approval for Breath Alcohol Screening Devices v1.0 using liquid simulators In-house method T408 based on the Home Office "Guide to Type Approval for Breath Alcohol Screening Devices v1.0
EVIDENTIAL BREATH ALCOHOL ANALYSIS INSTRUMENTS	Performance Tests - Moist ethanol vapour Evaluation of general functionality testing relating to devices for Type-Approval. Performance Tests - <i>In vivo</i> testing	In-house method T403 based on the Home Office "Guide to Type Approval for Evidential Breath Alcohol Analysis Instruments v2 using liquid simulators In-house method T409 based on the Home Office "Guide to Type Approval for Evidential Breath Alcohol Analysis Instruments v2
AQUEOUS ETHANOL SOLUTION	Quantification of Ethanol	In-house method T405 using HS GC-FID.
END		