


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>4668</b></p> <p>Accredited to ISO/IEC 17025:2017</p>	<b>Asbestos Laboratory Services (UK) Ltd</b>	
	<b>Issue No: 009    Issue date: 01 February 2022</b>	
<b>Ground Floor</b> Sycamore Court North Leigh Business Park North Leigh Witney Oxfordshire OX29 6SW	<b>Contact: Mr R Pearce</b> <b>Tel: +44 (0)1993 868636</b> <b>E-Mail: asbestoslabs@amiantus.co.uk</b> <b>Website: www.asbestoslabs.co.uk</b>	
<b>Testing performed by the Organisation at the locations specified</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b>                      Ground Floor                      Sycamore Court                      North Leigh Business Park                      North Leigh                      Witney                      Oxfordshire                      OX29 6SW</p> <p><b>Local contact</b>                      Mr Roy Pearce                      Tel: +44 (0)1993 868636                      E-Mail:                      asbestoslabs@amiantus.co.uk                      Website: www.asbestoslabs.co.uk</p>	Health and Hygiene Head Office  Asbestos – All Support Functions	A
<p><b>Address</b>                      Office 1                      24 Liverpool Gardens                      Worthing                      West Sussex                      BN11 1RY</p> <p><b>Local contact</b>                      Ms Olivia Pearce                      Tel: +44 (0)1903 211444                      E-Mail:                      asbestoslabs@amiantus.co.uk                      Website: www.asbestoslabs.co.uk</p>	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	B



4668

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Asbestos Laboratory Services (UK) Ltd**  
Issue No: 009 Issue date: 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<u>Health and Hygiene</u>  Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021  Documented In-House Method TP1 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, B
END			