
## **Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service** 

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



## DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
	Health and Hygiene	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method TP1 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248
END		