


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS TESTING 6093</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Allium Environmental Limited</p> <p>Issue No: 005 Issue date: 27 September 2022</p>	
	<p>Baldhu House Wheal Jane Earth Science Park Baldhu Truro TR3 6EH</p>	<p>Contact: Hadyn Veale Tel: 01872 276375 E-Mail: hadynveale@allium.uk.net Website: www.allium.uk.net</p>
<p>Testing performed by the Organisation at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address</p> <p>Wheal Jane Earth Science Park Baldhu Truro TR3 6EH</p> <p>Local contact</p> <p>Hadyn Veale Tel: 01872 276375 E-Mail: hadynveale@allium.uk.net</p>	<p>Health and Hygiene Head Office</p> <p>Asbestos – All Support Functions</p> <ul style="list-style-type: none"> • Quality System • Quality Audit Administration • Contract Review • Reporting 	A
<p>Address</p> <p>6 Testwood Lane Totton Southampton SO40 3AQ</p> <p>Local contact</p> <p>Hadyn Veale Tel: 01872 276375 E-Mail: hadynveale@allium.uk.net</p>	<p>Health and Hygiene</p> <p>Asbestos – All Support Functions</p> <ul style="list-style-type: none"> • Reporting 	D

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



6093
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Allium Environmental Limited

Issue No: 005 Issue date: 27 September 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method AT017 based on HSG 248	A, B, C
	Fibre counting	Documented In-House Method AT018, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method AT019, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method No TP1 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, D

END