

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 6376 Accredited to ISO/IEC 17025:2017	McHale Contracts & Plant Environmental LLP (trading as MCP Environmental)	
	Issue No: 005 Issue date: 31 March 2021	
	Enviro House Spartan Road South Bradford Trading Estate Low Moor Bradford BD12 0RY	Contact: Mr P Bates Tel: +44 (0)1274 601021 E-Mail: pbates@mcp-environmental.co.uk Website: www.mcp-environmental.com
Testing performed by the Organisation at the locations specified below		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<u>Health and Hygiene</u> Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248) Documented In-House Method No. LPR01 using stereo-microscopy, polarised light microscopy and dispersion staining based upon HSG 248
END		