


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 6536 Accredited to ISO/IEC 17025:2017	Clearview Environmental Ltd	
	Issue No: 006 Issue date: 16 March 2022	
	North Street Wigston Leicester Leicestershire LE18 1PS	Contact: Mr D Whitaker Tel: +44 (0)116 288 8256 E-Mail: david.whitaker@clearviewenv.com Website: www.clearviewenv.com
Testing performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address North Street Wigston Leicester Leicestershire LE18 1PS	Local contact Mr D Whitaker Tel: +44 (0)116 288 8256 E-Mail: david.whitaker@clearviewenv.com	Health and Hygiene Head Office Asbestos – All Support Functions
		A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



6536

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Clearview Environmental Ltd
Issue No: 006 **Issue date:** 16 March 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method SPM 003, based on HSG 248	A, B
	Fibre counting	Documented In-House Method SPM 003, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method SPM 003, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for asbestos identification	Documented In-House Method (SPM 001 Procedures for Surveying & Bulk sampling)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method SPM 002, using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			