


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING 6934</p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<p><b>Omega Asbestos Consulting Limited</b></p> <p>Issue No: 011 Issue date: 01 February 2022</p>	
	<p><b>First Floor</b> Wellsprings Business Centre Durham Rd West Bowburn Durham DH6 5AU</p>	<p><b>Contact: Mr G McKendrick</b> Tel: +44 (0)191 3770900 E-Mail: gary@omega-asbestos.co.uk Website: www.omega-asbestos.co.uk</p>
<p><b>Testing performed by the Organisation at the locations specified below</b></p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<p><b>Address</b> First Floor Wellsprings Business Centre Durham Road West Bowburn Durham DH6 5AU</p>	<p><b>Local contact</b> Gary McKendrick  Tel: +44 (0)191 3770900  Email: gary@omega-asbestos.co.uk Website: www.omega-asbestos.co.uk</p>	<p>Health and Hygiene Head Office</p> <p>Asbestos – All Support Functions</p>	A
<p><b>Address</b> AYE House, Admiralty Park Dunfermline Fife KY11 2YW</p>	<p><b>Local contact</b> Luke Stevenson  Tel: +44 (0)1383 630 633  E-Mail: luke@omega-asbestos.co.uk Website: www.omega-asbestos.co.uk</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• Quality Audit</li> <li>• Administration</li> <li>• Contract Review</li> <li>• Scheduling</li> <li>• Personnel</li> <li>• Equipment</li> <li>• Measurement</li> <li>• Traceability</li> <li>• Reporting</li> </ul>	C

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



6934  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Omega Asbestos Consulting Limited**  
Issue No: 011 Issue date: 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method Procedure Note 21 based on HSG 248	A, B & C
	Fibre counting	Documented In-House Method Procedure Note 21, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B & C
	4 Stage Clearance Process	Documented In-House Method Procedure Note 21, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B & C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method Procedure Note 20 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A & C
END			