


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 6934 Accredited to ISO/IEC 17025:2017	Omega Asbestos Consulting Limited Issue No: 013 Issue date: 26 May 2023	
	First Floor Wellsprings Business Centre Durham Rd West Bowburn Durham DH6 5AU	Contact: Mr G McKendrick Tel: +44 (0)191 377 0900 E-Mail: gary@omega-asbestos.co.uk Website: www.omega-asbestos.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address First Floor Wellsprings Business Centre Durham Road West Bowburn Durham DH6 5AU	Local contact Gary McKendrick Tel: +44 (0)191 3770900 Email: gary@omega-asbestos.co.uk Website: www.omega-asbestos.co.uk	Health and Hygiene Head Office Asbestos – All Support Functions	A
Address Suite 14 Ripponden Business Park Oldham Road Ripponden West Yorkshire HX6 4FF	Local contact Lee Connor Tel: +44 (0)1422 340 340 E-Mail: leeconnor@omega-asbestos.co.uk Website: www.omega-asbestos.co.uk	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none">• <i>Quality Audit</i>• <i>Administration</i>• <i>Contract Review</i>• <i>Scheduling</i>• <i>Personnel</i>• <i>Equipment</i>• <i>Measurement</i>• <i>Traceability</i>• <i>Reporting</i>	C

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



6934
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Omega Asbestos Consulting Limited
Issue No: 013 Issue date: 26 May 2023

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method Procedure Note 21 based on HSG 248	A, B & C
	Fibre counting	Documented In-House Method Procedure Note 21, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B & C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method Procedure Note 21, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B & C
	Sampling of bulk materials for asbestos identification	Documented In-House Method 01 Procedure note 01 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method Procedure Note 20 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A & C
END			