


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS TESTING 7687</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<p><b>RPS Occupational Health Limited</b></p> <p>Issue No: 004    Issue date: 09 August 2019</p>	
	<p><b>Innospec Manufacturing Park</b> Oil Sites Road Ellesmere Port CH65 4EY United Kingdom</p>	<p><b>Contact: Mr Francis McTeigue</b> Tel: +44 (0) 151 348 5678 E-Mail: <a href="mailto:rpshiblab@rpsgroup.com">rpshiblab@rpsgroup.com</a></p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
BODY FLUID	<u>Chemical Tests and Workplace Testing</u>	Documented in-house methods:
Human blood	Lead levels	L/Q/01 using graphite furnace atomic absorption spectroscopy
Human urine	Creatinine (0-20mg/dl)	L/Q/04 using enzyme immunoassay followed by spectrophotometric detection
	pH (3-11)	L/Q/04 using colourimetry
	Specific gravity (1.001-1.035)	L/Q/04 using colourimetry
	Presumptive screening for the presence of the following drugs (cut-off limit): Amphetamines (500ng/ml) Benzodiazepines (200ng/ml) Barbiturates (200ng/ml) Cannabinoids (50ng/mL) Cocaine as Benzoyllecgonine (150ng/mL) Methadone (300ng/ml) Phencyclidine (25ng/ml) Propoxyphene (300ng/ml) Opiates (300ng/ml) Ethanol (107ng/ml)	L/Q/04 using enzyme immunoassay followed by spectrophotometric detection
<p>END</p>		