

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 7691 Accredited to ISO/IEC 17025:2005	North Star Environmental Limited	
	Issue No: 004	Issue date: 23 November 2016
	Suite A1 Beech House Oaklands Office Park Hooton CH66 7NZ	Contact: Mr Karak Burns Tel: +44 (0)151 538 3141 E-Mail: info@northstarenvironmental.co.uk Website: www.northstarenvironmental.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Suite A1 Beech House Oaklands Office Park Hooton CH66 7NZ Local contact Mr Karak Burns Tel: +44 (0)151 538 3141 E-Mail: info@northstarenvironmental.co.uk Website: www.northstarenvironmental.co.uk	Health and Hygiene	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



7691

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

North Star Environmental Limited
Issue No: 004 Issue date: 23 November 2016

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248:February 2005 Membrane Filter Method using Phase Contrast Microscopy (PCM) (Documented In-house Procedure NSTM2)	A, B, C
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure NSTM3)	B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure NSTM3)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 Documented in-House Procedure NSTM4_02 Asbestos in Bulk Materials using stereo microscopy, polarised light microscopy, dispersion staining and phase contrast	A
END			