


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |   |  |
|--|---|--|
| <br><b>7783</b><br><br>Accredited to<br><b>ISO/IEC 17025:2017</b> | <b>Davidson and Hardy (Laboratory Supplies) Limited</b> |  |
|  | Issue No: 002    Issue date: 09 April 2020              |  |
|  | 453 – 455 Antrim Road<br>Belfast<br>BT15 3BL            | Contact: Mr J Featherstone<br>Tel: +44 (0)7501 255405<br>Fax: +44 (0)28 9037 0115<br>E-Mail: joe@dhlab.com<br>Website: www.dhlab.com |
| Testing performed by the Organisation at the locations specified below   |   |  |

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

| Location details        | Activity                           | Location code |
|-------------------------|------------------------------------|---------------|
| Customer Premises (any) | Testing of Controlled Environments | Any           |



7783  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Davidson and Hardy (Laboratory Supplies) Ltd**  
Issue No: 002 Issue date: 09 April 2020

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Materials/Products tested  | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used   | Location Code |
|--|---|--|---------------|
| <b>Controlled Environments</b><br>Including:<br>Microbiological safety cabinets<br>Class I and II<br>Laminar airflow workstations<br>Isolators<br>Fume Cupboards & Hoods | Airflow velocity<br>(0.25 m/s to 1.5 m/s)             | In-house documented methods based on the relevant sections of the following standards:<br><br>Microbiological safety cabinets:-<br>EN 12469:2000<br><br>Fume Cupboards:<br>EN 14175-3-2004 | Any           |
| END  |   |  |               |