

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING</p> <p>7830</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p><b>Northburn Oils Ltd</b></p> <p>Issue No: 012    Issue date: 06 September 2024</p>	
	<p>NOL Testing Laboratory 70 Northburn Road Coatbridge Scotland ML5 2HY</p>	<p>Contact: Mr P Morrison Tel: +44 (0)1236 427514 E-Mail: <a href="mailto:peterm@northburn.co.uk">peterm@northburn.co.uk</a></p>
<p>Testing performed at the above address only</p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PROCESSED FUEL OIL (PFO) and PFO DISTILLATE	<u>Chemical and Physical Tests</u>	
	Carbon Residue	IP 398/15 BS 2000-398:14 BS EN ISO 10370:14 Micro method
	Flashpoint	IP 523/23 BS 2000-523:23 BS EN ISO 3679:22 Rapid equilibrium closed cup
	Total Sediment	IP 375/11 BS 2000-375:11 ISO 10307-1:09
	Strong Acid Number (Qualitative Test)	IP 139/98(17) BS 2000-139:98 ISO 6618:96
	Sulphated Ash	IP 550/08
	Water content	IP 74/2000(14) BS 2000-74:00 ISO 3733:99
	<u>Sampling</u>	
	Manual sampling	Documented in-house method WIN04 based on: IP 475/05 BS 2000-475:04 BS EN ISO 3170:04
Sulphur content	ASTM D2622-21 using in house method LABWIN09 and WDXRF	



7830

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

#### Northburn Oils Ltd

Issue No: 012 Issue date: 06 September 2024

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PROCESSED FUEL OIL (PFO) and PFO DISTILLATE	Chemical and Physical Tests  Determination of Trace Metals: Pb, Ni, Cr, Cu, Zn, As, Cd, Tl, Sb, Co, Mn, and V  Mercury  Total Halogens as Chlorine	IP 593/11 using LABWIN10 and WDXRF  IP 608/15 using LABWIN15 and WDXRF  IP 503/04 using LABWIN11 and WDXRF
END		