


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>Accredited to ISO/IEC 17025:2017</b>	<b>Casa Environmental Services Ltd</b>  <b>Issue No: 008    Issue date: 29 May 2024</b>	
	<b>The Haybarn Londonderry Farm Keynsham Road Willsbridge Bristol BS30 6EL United Kingdom</b>	<b>Contact: Mr Ben House Tel: +44 (0) 1179 322 323 E-Mail: <a href="mailto:benhouse@casaenvironmental.co.uk">benhouse@casaenvironmental.co.uk</a> Website: <a href="http://www.casaenvironmental.co.uk">www.casaenvironmental.co.uk</a></b>

**Testing performed by the Organisation at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<b>Address</b>  The Haybarn Londonderry Farm Keynsham Road Willsbridge Bristol BS30 6EL United Kingdom	<b>Local contact</b>  Contact: Mr Ben House Tel: +44 (0) 1179 322 323 E-Mail: <a href="mailto:benhouse@casaenvironmental.co.uk">benhouse@casaenvironmental.co.uk</a> Website: <a href="http://www.casaenvironmental.co.uk">www.casaenvironmental.co.uk</a>	Health and Hygiene  Asbestos – All Support Functions	A
<b>Address</b>  Units 110 (Office) / 313 (Lab) Solent Business Centre Millbrook Road West Southampton S015 0HW	<b>Local contact</b>  Contact: Mr Ben House Tel: +44 (0) 1179 322 323 E-Mail: <a href="mailto:benhouse@casaenvironmental.co.uk">benhouse@casaenvironmental.co.uk</a> Website: <a href="http://www.casaenvironmental.co.uk">www.casaenvironmental.co.uk</a>	Health and Hygiene	B

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	C
Mobile laboratories	Health and Hygiene	D



7914  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Casa Environmental Services Ltd**

**Issue No: 008 Issue date: 29 May 2024**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method TP04 based on HSG 248	C
	Fibre counting	Documented In-House Method TP04, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, C, D
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method TP04, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	C
	Sampling of bulk materials for asbestos identification	Documented In-House Method TP01 based on HSG 248	C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method TP03 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, B
END			