


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>Accredited to ISO/IEC 17025: 2017</b>	<b>Essex Asbestos Testing Ltd</b>	
	<b>Issue No:</b> 005 <b>Issue date:</b> 01 February 2022	
	<b>15 Lampits Lane Corringham SS17 9AD</b>	<b>Contact: Mr I Donaldson Tel: +44 (0) 1375 643 863 E-Mail: <a href="mailto:info@essex-asbestos-testing.co.uk">info@essex-asbestos-testing.co.uk</a> Website: <a href="http://www.essex-asbestos-testing.co.uk">www.essex-asbestos-testing.co.uk</a></b>
<b>Testing performed by the Organisation at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<b>Address</b> 15 Lampits Lane Corringham SS17 9AD	<b>Local contact</b> Mr I Donaldson  Tel: +44 (0) 1375 643 863	Health and Hygiene  Asbestos – All Support Functions	A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



8007  
Accredited to  
ISO/IEC 17025: 2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Essex Asbestos Testing Ltd**  
**Issue No:** 005 **Issue date:** 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method TP03 based on HSG 248	A, B
	Fibre counting	Documented In-House Method TP03, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
END			