


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 8229 Accredited to ISO/IEC 17025:2017	Index Microbiology Limited Issue No: 015 Issue date: 18 December 2024	
	Index Microbiology The Churchill Building University of Bristol Langford BS40 5DU United Kingdom	Contact: Dr Robert Morley Tel: +44 (0)117 428 3553 E-Mail: info@indexmicrobiology.com Website: www.indexmicrobiology.com
	Testing performed at the above address only	

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>WATERS</p> <p>Non-Regulatory Drinking Water, Process Water, Recreational Water including Pool Water</p> <p>Non-Regulatory Drinking Water, Process Water (including Cooling Towers), Recreational Water including Pool Water</p> <p>Non-regulatory Drinking Water, Process Waters, Recreational Water including Pool Water</p> <p>Non-Regulatory Drinking Water, Surface Water, Recreational Water including Pool Water</p> <p>Drinking Water (mains), Surface Water</p> <p>Purified waters (endoscope rinse waters)</p> <p>Purified waters (endoscope rinse water and RO water) and hospital tap water</p> <p>Purified waters (endoscope rinse water, water for Haemodialysis)</p>	<p><u>Microbiological Tests</u></p>	
	<p>Enumeration of <i>Pseudomonas aeruginosa</i></p>	<p>IM01 -by filtration (based on Microbiology of Drinking Water Pt 8b:2015)</p>
	<p>Detection and Enumeration of <i>Legionella</i> species including identification of <i>L. pneumophila</i> serogroups 1-14</p>	<p>IM02 – in-house method using membrane filtration and GVPC agar</p>
	<p>Colony count at 22°C for 3 days & 37°C for 1 & 2 days</p>	<p>IM04 - by pour plate (Microbiology of Drinking Water Pt 7:2020)</p>
	<p>Confirmed total coliforms and <i>Escherichia coli</i> by IDEXX Coliort</p>	<p>IM05 - IDEXX Quantitray (based on Microbiology of Drinking Water Pt 4D 2016) MPN</p>
	<p>Enumeration of <i>Enterococci</i></p>	<p>IM015- IDEXX Enterolert (Microbiology of drinking water Pt 5 2012) MPN</p>
	<p>Enumeration of Environmental Mycobacterium 30°C for 28 days</p>	<p>IM030 by membrane filtration and microscopy/ZN stain based on Health Technical Memorandum 01-06: Decontamination of Flexible Endoscopes - Part E, Testing Methods</p>
	<p>Enumeration of <i>Pseudomonas aeruginosa</i></p>	<p>IM01 - by filtration based on Microbiology of Drinking Water Pt 8b:2015 and HTM 01-06 Part E 2016</p>
	<p>Colony count at 22°C 7 days, 30°C for 5 days</p>	<p>IM03 - membrane filtration based on Health Technical Memorandum 01-06: Decontamination of Flexible Endoscopes - Part E Testing Methods BS EN 23500-3:2024</p>



8229
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Index Microbiology Limited
Issue No: 015 Issue date: 18 December 2024

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATERS Purified waters (endoscope rinse water, water for Haemodialysis) Liquid products for household and personal wipes Cosmetic Wipes	<u>Microbiological Tests</u> Colony count at 22°C 7 days, Enumeration of aerobic mesophilic bacteria at 22°C for 3 days & 37°C for 2 days Enumeration of aerobic mesophilic bacteria, 37°C for 2 days and yeasts and moulds at 22°C for 3 days	IM04 – by pour plate (Microbiology of Drinking Water Pt 7:2020) and BS EN 23500-3:2019 IM06 –(In house method) IM06 –(In house method)
END		