

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING</p> <p>8929</p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<h3>True Refrigeration UK Limited</h3> <p><b>Issue No:</b> 007      <b>Issue date:</b> 01 November 2024</p>	
	<p><b>Field End Road</b> <b>Goldthorpe</b> <b>Rotherham</b> <b>S63 9EU</b></p>	<p><b>Contact: Michael Russell</b> <b>Tel: +44 (0)1709 888 888</b> <b>E-Mail: <a href="mailto:mrussell@truemfg.com">mrussell@truemfg.com</a></b> <b>Website: <a href="http://www.truemfg.com">www.truemfg.com</a></b></p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>Refrigerated storage cabinets and counters for professional use.</p> <p>Refrigerated display cabinets</p> <p>Commercial beverage coolers- classification requirements and test conditions</p> <p>Energy consumption of vending machines</p>	<p><b>Performance and Energy Tests</b></p> <p>Product Temperature Range</p> <p>-20°C to 0°C -1°C to 10°C</p> <p>Climatic Conditions</p> <p>Temperature 25°C to 40°C Tolerance ± 1°C</p> <p>Relative Humidity 40% - 60% Tolerance ±3 to ±5 units</p> <p>Energy consumption EEI, SEC and TEC values</p>	<p>BS EN 16825:2016 (Withdrawn) BS EN ISO 22041:2019</p> <p>BS EN 23953-1: 2023 BS EN 23953-2: 2023 BS EN 23953-1:2015 (withdrawn) BS EN 23953-2:2015 (withdrawn)</p> <p>BS EN ISO 22044:2022 BS EN IEC 63252:2020</p>
<p>Enclosures and cabinets (all types)</p>	<p><b>Ingress Protection</b></p> <p>IP2X (Protection against solid foreign objects of ≥12.5mm Ø )</p> <p>IP3X (Protection against solid foreign objects of ≥ 2.5mm Ø )</p> <p>IP4X (Protection against solid foreign objects of ≥1mm Ø )</p> <p>IPX1 (Protection against dripping water)</p>	<p>EN 60529:1992 + A2 :2013</p>
<p>END</p>		