

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 Accredited to ISO/IEC 17025:2017	Not One Iota Ltd Trading as Iota	
	Issue No: 012 Issue date: 27 August 2024	
	10 Tandragee Road Portadown County Armagh BT62 3BQ United Kingdom	Contact: Kate Gorman Tel: +44 (0)2838 525025 E-Mail: kate@notoneiota.org Website: www.notoneiota.org
Testing performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 10 Tandragee Road Portadown Co. Armagh BT62 3BQ United Kingdom Kate Gorman Tel: +44 (0)2838 525025 Email: kate@notoneiota.org	Health and Hygiene Asbestos – All Support Functions Microbiology testing	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Not One Iota Ltd
Trading as Iota

Issue No: 012 **Issue date:** 27 August 2024

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method SOP-12, based on HSG 248	B
	Fibre counting	Documented In-House Method SOP-12, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
	4 Stage Clearance Process	Documented In-House Method SOP-12, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method SOP-11, using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
WATERS as specified	<u>Microbiological tests</u>	Documented in-house methods	
Drinking water (non-regulatory), Recreational water	Enumeration of: Total aerobic colony count at : 22°C 37°C	Method SOP03 based upon Microbiology of Drinking Water part 7, 2020	A
	<u>Coliforms and <i>Escherichia coli</i> (confirmed)</u>	Method SOP04 by membrane filtration using MLGA based on Drinking Water part 4B, 2016	A
	<u><i>Pseudomonas aeruginosa</i> (confirmed)</u>	Method SOP05 by membrane filtration based on Drinking Water part 8, 2015	A



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Not One Iota Ltd
Trading as Iota

Issue No: 012 **Issue date:** 27 August 2024

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS as specified Drinking water (non-regulatory), Recreational water, Process and Healthcare waters	<u>Microbiological tests</u> (cont'd) Enumeration and identification of: <i>Legionella</i> spp. and <i>Legionella pneumophila</i> serogroups 1-14	Documented in-house methods Method SOP02 based on ISO11731:2017 using GVPC, with Identification using Oxoid latex agglutination kit (Matrix A & B: Procedure 8, 9 & 10, Media C)	A
END			