


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>9192</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<p><b>Perfectus Biomed Limited</b></p> <p>Issue No: 004    Issue date: 06 November 2018</p>	
	<p>Tech Space One Sci Tech Daresbury Keckwick Lane Cheshire WA4 4AB</p>	<p>Contact: Dr Samantha Westgate Tel: +44 (0)1925 737237 E-Mail: sam@perfectusbiomed.com Website: www.perfectusbiomed.com</p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>DISINFECTANTS, ANTISEPTICS, SOAP and DETERGENTS</p> <p>Disinfectants, antiseptics, biocides, gels, topical ointments, and wound dressings materials</p>	<p>Disinfectant Efficacy Testing of Biofilms for specified organisms:</p> <p><i>Pseudomonas aeruginosa</i></p>	<p>Documented in-house methods</p> <p>1) SOP 536 test method for testing disinfectant efficacy against biofilm grown according to ASTM E2799-17 using Minimum Biofilm Eradication Concentration (MBEC) assay at 37°C for 24 or 72h</p> <p>2) SOP 537 test method for testing disinfectant efficacy against biofilm grown according to ASTM E2562-17 using high shear and continuous flow using CDC biofilm reactor at 37°C for 24 or 72h</p> <p>3) SOP 538 test method for testing disinfectant efficacy against biofilm grown according to ASTM E2647-13 using drip flow biofilm reactor with low shear and continuous flow at 37°C for 24 or 72h</p>



Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Perfectus Biomed Limited**  
**Issue No: 004 Issue date: 06 November 2018**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>DISINFECTANTS, ANTISEPTICS, SOAP and DETERGENTS (cont'd)</p> <p>Disinfectants, antiseptics, biocides, gels, topical ointments, and wound dressings materials (cont'd)</p>	<p>Disinfectant Efficacy Testing of Biofilms for specified organisms: (cont'd)</p> <p><i>Staphylococcus aureus</i></p> <p><i>Candida albicans</i></p>	<p>Documented in-house methods</p> <p>1)SOP 536 test method for testing disinfectant efficacy against biofilm grown following the principles of ASTM E2799-17 using Minimum Biofilm Eradication Concentration (MBEC) assay at 37°C for 24 or 72h</p> <p>2)SOP 537 test method for testing disinfectant efficacy against biofilm grown following the principles of ASTM E2562-17 using high shear and continuous flow using CDC biofilm reactor at 37°C for 24 or 72h</p> <p>3)SOP 538 test method for testing disinfectant efficacy against biofilm grown following the principles of ASTM E2647-13 using drip flow biofilm reactor with low shear and continuous flow at 37°C for 24 or 72h</p> <p>SOP555 test method for testing disinfectant efficacy against biofilm grown following the principles of ASTM E2562-17 using high shear and continuous flow using CDC biofilm reactor at 37°C for 72h</p>
<p>END</p>		