


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Isopharm Limited</h3> <p>Issue No: 011    Issue date: 24 January 2020</p>	
	<p>79 Leigh Street Sheffield S9 2PR</p>	<p>Contact: Adam Stout Tel: +44 (0)114 354 0136 E-Mail: adam.stout@isopharm.co.uk Website: www.isopharm.co.uk</p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>WATERS</p> <p>Endoscopy Washer-Disinfector Final Rinse Waters (Mains fed and Reverse Osmosis), Reverse Osmosis Water and Potable Water</p>	<p><u>Chemistry Tests</u></p> <p>Total Organic Carbon</p> <p>Specified analytes:</p> <ul style="list-style-type: none"> <li>• Phosphate</li> <li>• Silicate</li> <li>• Total Hardness</li> <li>• Sulphate</li> <li>• Ammonia</li> <li>• Chloride</li> <li>• Nitrate (by calculation)</li> <li>• Total Oxidised Nitrogen</li> <li>• Nitrite</li> </ul> <p>pH and Conductivity</p> <p>Total Dissolved Solids</p> <p><u>Microbiology Tests</u></p> <p>Total Aerobic Colony Count</p>	<p>Documented in-House Methods:</p> <p>AP4 based on ISO 8245:1999 using TOC Analyser</p> <p>AP2 based on ISO 15923-2:2017 using Discrete analyser</p> <p>AP5 based on ISO 10523:2012 and BS EN 27888:1993 using robotic probe manipulation and Consort meter</p> <p>AP8 by evaporation and gravimetry based on Suspended, Settleable and Total Dissolved Solids. Methods for the Examination of Waters and Associated Materials, 1980 to meet the requirements of HTM 01-01 Part D, 2016</p> <p>AP1.3 using membrane filtration onto TSA, incubated at 35°C for 3 days, based on HTM 2030 1997 (withdrawn)</p>



9508

Accredited to  
ISO/IEC 17025:2017

### Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Isopharm Limited

Issue No: 011 Issue date: 24 January 2020

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>WATERS (cont'd)</p> <p>Endoscopy Washer-Disinfector Final Rinse Waters (Mains fed and Reverse Osmosis), Reverse Osmosis Water and Potable Water</p> <p>Endoscope Surrogate Devices and Lumened Medical Devices</p>	<p><u>Microbiology Tests</u> (cont'd)</p> <p>Total Aerobic Colony Count</p> <p><i>Pseudomonas aeruginosa</i></p> <p>Environmental Mycobacteria</p> <p>Inoculation, recovery and enumeration of <i>Pseudomonas aeruginosa</i> and <i>Staphylococcus aureus</i> from lumened devices</p>	<p>Documented in-House Methods:</p> <p>AP1.5 using membrane filtration onto R2A, incubated at 30°C for 5 days, based on BS EN ISO 15883-1:2009 + A1: 2014, HTM 01-06 Part E; 2016, WHTM 01-06 and CFPP 01-06 (withdrawn)</p> <p>AP13 using membrane filtration based on Microbiology of Drinking Water Part 8, 2015 to meet the requirements of HTM 01-06 Part E, 2016</p> <p>AP3 using membrane filtration based on HTM 01-06 Part E, 2016, confirmed through acid-fast bacilli staining</p> <p>AP11 using inoculation technique, elution by flushing and plating out on TSA based on BS EN ISO 15883-4:2009 6.6.2 and Annex B, plus HTM 01-06 Part E 2016</p>
END		